ULTRA-CONTROLO

Anaesthetic Gas Scavenging

Active AGSS Unit, 1L - EN ISO 9170-2

TERMINAL UNITS

Anaesthetic Gas Scavenging

Active AGSS Unit, 1L - EN ISO 9170-2



Active AGSS terminal unit according to EN ISO 9170-2, the 1L type socket is the final part of the system and the point at which the operator operates, through the relevant connections, connections and disconnections to evacuate the anesthetic gases.

Features

- Venturi-driven with compressed air
- Green suction indicator
- Containment box and fixing plate
- Free, external or recessed installation

Standards

- UNI EN ISO 7396-2
- UNI 9507
- UNI EN ISO 9170-2
- UNI EN ISO 15001

Maintenance kits

- AGSS · Flux indication / green viewer Ø 6 replacement

Components

- Body and ejector in brass
- Attuator in anodised aluminium
- Aluminium base
- Anodised aluminium release gear
- Stainless steel springs
- EPDM PEROX O-Ring
- ABS box
- Inox steel cover

GAS	DESCRIPTION	TYPE	FLOWRATE AT 4.5 BAR	PRESSURE DROP AT 90 L/MIN	WEIGHT
AGSS	T.U. AGSS Venturi ISO 9170-2 - free installation	1L	50 l/min	0.035 bar	0.51 kg
AGSS	T.U. AGSS Venturi ISO 9170-2 - external	1L	50 l/min	0.035 bar	0.89 kg
AGSS	T.U. AGSS Venturi ISO 9170-2 - recessed	1L	50 l/min	0.035 bar	1.03 kg